

OCT 29 2010

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130 289 ✓
Name		James Kimura	ID# 5677 ✓ Date 10/23/10 ✓
A	Agency	Alaska State Troopers	Phone # 225-5118
Instrument Location Ketchikan Post			
Alco S/N		79268 ✓	Target Value .081 ✓ High Pressure 900 PSI
B	Alco Test Values		
	1st Alco	.081 ✓	2nd Alco .080 ✓
Signature <u>[Signature]</u>			
(OVER)			

BMB
12/9/10

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of DEC, 2010.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130289

BMB
12/9/10

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

OCTOBER 23, 2010

OPERATOR'S NAME:

KIMURA/JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 000000

DEPT/AGENCY: KETO

CASE/REPORT: 10-00000

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: 78268

— BREATH ANALYSIS —

.081 ADJUSTED FOR 29.20 in
ALCO TARGET .079 23:32
BLANK TEST .000 23:33
INTERNAL STANDARD VERIFIED 23:33
ALCO TV 29.20 in .081 23:33
BLANK TEST .000 23:34
SUBJECT SAMPLE .000 23:34
BLANK TEST .000 23:35
ALCO TV 29.20 in .000 23:35
BLANK TEST .000 23:36

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

OCTOBER 23, 2010

TIME 23:29

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.20 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~